
KUSAM-MECO®

AN ISO 9001:2000 COMPANY



DIGITAL ANALOG STORAGE OSCILLOSCOPE

MODEL KM 20-5030 / KM 20-5060

INSTRUCTION MANUAL

CONTENTS

Title	Page No.
1. Features	1
2. Accessories	1
3. Precautions	2
4. Controls And Connectors	5
5. How To Produce the Bright Line	12
6. Method For Connecting Signals	14
7. Measuring Procedure	16
8. Adjustments	26
9. Maintenance	27
10. Specifications	27
11. External View	31
12. Test Certificate	32
13. Warranty	33

1. FEATURES

The V-5060/V552 is a portable-type, advanced-class oscilloscope with a bandwidth of DC to 60 MHz /50MHz designed with the emphasis on operability and portability and has a following features.

(1) Wide bandwidth

The instrument has a bandwidth from DC to60 MHz 50MHz

(2) High sensitivity

Sensitivity is1 mV/div

(3) Large 6 " screen:

Employment of a large square CRT makes waveforms easier to observed

(4) internal graticule:

Employment of an internal graticule CRT permits waveforms observation to be made without parallax error

(5) DC OFFSET:

Allows a proportionate observation of any part of an input waveform even when it is in a large amplitude.

(6) ALTMAG

Allows simultaneous observation of X1and X10 sweeping waveforms

(7) ALT TRIG

Even an observation of two waveforms of different frequencies, the waveform of the each cannel is stably triggered.

(8) TV synchronization

Employment of a new TV sync separator circuit allows the instrument to observe TV signals stably.

(9) Auto focusing

Focusing shift is automatically corrected.

2.ACCESSORIES

This instrument is shipped with the following standard accessories.

2 Probes

V-5060: MEI LI (1:10)

V-552 : MEI LI (:10)

1 AC power supply coad

1 Operation manual

In case of(EM) TYPE : Without Probes

3.PRECAUTIONS

Precautions to be observed to lengthen the service life of this instrument.

Installation site


* Avoid installing instrument in an extremely hot or cold place
  - Avoid placing this instrument in a place exposed to sun-light for a long period of time, in a closed car in mid-summer ,or near a room heating device such as a stove
  - The operating maximum ambient temperature is+40℃
* Do not use instrument that has been left outdoors on a cold winter day.
  The operating ambient temperature is 0℃ or more



* Avoid moving the instrument rapidly from a hot place to a cold place of vice versa, or condensation may form on
* Keep the instrument away from damp air, inside of the instrument water and dust.
  Unexpected trouble may be caused when the instrument is placed in a damp or dusty place.
  The operating ambient humidity is35-85%
  Since an accidental intrusion of water may also cause troubles, do not place a water -filled containers such as a vase the oscilloscope



* Do not place the instrument in a place where vibration is strong. Avoid using the instrument at a place vibrating violently. Since the oscilloscope is a precision instrument, excessively strong vibrations may cause damage.
* Do not place the instrument near a magnet or magnetic body. An oscilloscope is an equipment using electron beam. Therefore, do not bring a magnet close to the instrument or do not use the instrument near an equipment generatoing strong magnetic force.



Handling

* Do not put a heavy objects on the oscilloscope.
  Do not block the ventilation holes.

* Do not apply a heavy shock to the oscilloscope





* Do not insert a wire, pin, etc. through the ventilation hole.
  Do not drag the set, leaving the probe attached to it.





Do not leave a hot soldering iron on the cabinet or the screen.

Do not try to turn the instrument upside down. Otherwise, knobs may be broken





Do not use the instrument upright, leaving BNC cable connected to EXT BL ANKING terminal on the rear panel. Otherwise, the cable may be damaged.



When operation is faulty

Recheck the operating procedure and if problem persists contact a nearly service station of agent.

Care and repair

Removal of stain from the case

When the outside of the case is stained, remove the stain by first wiping it lightly with a cloth moistened with neutral washing agent and then wipe the surface with a dry cloth

Never use strongly volatile agent such as benzine and thinner.



- When the panel surface is stained, remove the stain in similar way with a clean a clean, soft cloth. When heavy stains are present, first remve the stains by wiping the surface lightly with a cloth moistened with diluted neutral washing agent or with alcohol and then wipe thoroughly with a dry cloth.
- When dust has accumulated on the inside, remove it by using dry brush, or by using the exhaust of a compressor or a vacuum cleaner. By a SERVICE PERSONNEL.

NOTE: When opening the case ,pull out the power supply plus beforehand without fail.
When cleaning the inisde, insure beforehand that no electricity remains in the condensers of the power supply circuit." NOT FOR AN OPERATOR"

Operation precautions

prior to shipment, the voltage selector is set property.
When the oscilloscope is intended to be used on a different.
Voltage, re locate the voltage selector in the procedures which follow.

1) Disconnect the power connector.
2) Insert a screwdriver into the right side or the cap and remove the fuse holder cap.
3) Mount the fuse holder cap in to the fuse holder so that the marking of the correct voltage faces up.
4) Connect the power connector.


* Do not increase the brightness too much.
Do not increase the brightness of the spot and trace too much your. Eyes may be strained and the fluorescent surface of CRT may be burnt.
* Do not apply an excessive voltage.
The input withstand voltage of each input connector and probe input is as follows. Never apply a voltage higher than specified.

INPUT direct,	300V (DC+AC peak at 1 kHz)
When×10 probe is used	400V (DC+AC peak at 1 kHz)
When×1 probe is used	300V (DC+AC peak at 1 kHz)
EXT TRIG INPUT	300V (DC+AC peak)
EXT BLANKING	30V (DC+AC peak)

Calibration Interval

To maintain instrument accuracy, performs the calibration of the V-5060 / V-552 at least every 1000 hours of operation, or every months if used infrequently.

4. CONTROLS AND CONNECTORS



① **POWER switch**

The POWER is set on at the pushed-in position, and set off at the released position

② **Power lamp**

This lamp goes on when the power supply is in ON state.

③ **FOCUS control**

After obtaining an appropriate brightness by operating INTENsity. adjust FOCUS until the bright line is clearest. Although the focus is also corrected automatically when INTEN is rotated, the focus is sometimes slightly shifted.

④ **SCALEILLUM control**

Controls graticule illumination Useful to illuminate the graticule when viewing in a dark area, photographing.

⑤ **TRACE ROTATION control**

Used to aline the trace of CRT with the horizontal graticule.

⑥ **INTEN sity control**

This knob also works as the brightness adjust variable resistor. Brightness is increased by rotating INTENsity clockwise

⑦ **FUSE**

⑧ **AC INLET**

This is inlet for detachable AC power cord.

(2) Controls of vertical deflection system

⑨ **CH1 INPUT connector**

BNC connector for vertical axis input.

The signal input ro this terminal becomes the X-axis signal when the instrument is used as an X-Y oscilloscope.

⑩ CH2 INPUT connector

The same as Ch1, but when the instrument is used as an X-Y oscilloscope, the signal input to this terminal becomes the Y-axis signal.

⑪⑫ **Input coupling switches(AC-GND-DC)**

The switches used to select the coupling system between the input signal and vertical axis amplifier.

AC A t this setting the signal is connected through a condenser. The DC component of the input signal is cut off and only the AC component is displayed.

GND At this setting the input to the vertical axis amplifier is grounded.

DC Ar this setting the input signal is directly connected to the vertical axis amplifier and displayed unchanged, including the DC component.

⑬⑭ **VOLTS/DIV select switches**

A step attenuator which selects vertical deflection factor. Set it to an easily observable range corresponding to the amplitude de of the input signal.

Multiply the reading by 10 when the 10:1 probe is used in combination with the instrument.

⑮⑯ **VAR PULL×5GAIN controls**

Fine tuning device used to vary the vertical deflection sensitivity continuously. Attenuation of less than 1/2.5 is obtained when this device rotated in the reverse direction of the arrow to the full.

This control is used when comparing waveforms or when measuring the rise time of a square wave in 2-channel observation. Normally this control is left rotated in the direction of the arrow to the full. When the knob is at PULL position(pulled up state)the gain, of the vertical axis is magnified 5tims and the maximum sensitivity becomes 1mV/DIV.

⑰⑱ **UNCAL lamp**

Light when VAR is out of CAL detent position

⑲ **POSITION control**

PULL DC OFFSET

This knob used to adjusting the position of the vertical axis.

The image rises with the clockwise rotation of this knob and falls with the counterclockwise rotation.

When the knob is pulled up, the adjustment range of the trace position of the vertical axis can be magnified by the DC OFFSET function. Therefore, the peak value of an input waveform with large amplitude can be measured. (keep pushed-in for the normal operation.)

⑳ **POSITON PULL INVERT control**

The same as CH1,but when the knob is at PULL position (pulled up state), this is used to inverse the polarity of the input signal applied to CH2.

This control is conveniently used in the comparison of two waveforms having different polarity or in the observation of the waveform of the difference signal (CH1)-(CH2)betwen CH1and CH2using ADD.

㉑ **MODE select switch**

This switch is used to select the operation mode of the

Vertical deflection system.

CH1 Only the signal that has been applied to CH1 appears on the screen.

CH2 Only the signal that has been applied to CH2ap-pears on the screen.

ALT Signals applied respectively to CH1 and CH2 appear on the screen alternatively at each sweep. This setting is used the sweep time is short in-2-channel observation.

CHOP At this setting the in put signals applied respecting to CH1 and CH2 are switched at about 250kHz independent of the sweep and at the same time appear on the screen, This setting is used when the sweep time is long in 2-channel observation.

ADD The algebraic sum of the input signals applied respectively to CH1 and CH2 appears on the screen.

㉒ **CH1 OUTPUT connector**

Output connector providing a sample of the signal applied to the CH1 connector.

㉓ **DC OFFSET VOLTOUT**

This is the output connector to retdout the voltage measurement with a digital multi-meter,etc.,when the instrument is set to the DC OFFSET mode.
(Except : ×5 GAIN, out of CAL)

㉔㉕ **DC BAL adjustment contrlos**

There are used for the ATT balance adjustment.
See .8 ADJUSTMENTS . . For the details.
Controls of Horizontal deflection system

㉖ **TIME/DIV select switch**

Sweep time ranges are 19 steps from 0.2 us/div to 0.2s/div.

X-Y This position is used when using the instrument as an X-Y oscilloscope.
In this position the X(horizontal)signal is connected to the input of CH1;the Y (vertical) signal is applied to the input of CH2 and has a deflection range from less than one millivolt to 5 volts /div at a reduced band -wide of500kHz.

㉗ **SWP VARiable control**

This control works as CAL and the sweep time is calibrated to the value indicated by TIME/DIV
TIME/DIV. Of sweep can be varied continuously when shaft is out of CAL position.
Then the control is rotated in the direction of arrow to the full, the CAL state is produced and the sweep time

is calibrated to the value indicated by TIME/DIV. Countertclockwise rotation to the full delays the sweep by 2.5times or more.

㉘ **Sweep UNCAL lamp**


Light when SWP VAR is out of CAL detent position.

㉙ **POSITION control**
PULL×10MAG

This knob is used to move the bright line in horizontal directions. It is indespensable in the measurement of the time of waveform.

Bright line is move toward right when the knob is rotated clockwise and toward left with counterclockwise rotation.

Sweep is magnified 10 times by pulling out knob of POSITION. In this case the sweep times is1/10of the value indicated by TIME/DIV.Bring the position of the waveform desired to be magnified observed to the outer of the scale by operating ◀ ▶ POSITION of the horizontal axis. Next, switch x 10MAG switch to PULL (pulled out state).Then the waveform placed at the center is magnified in right and left directions. The sweep time in this case is 10 times the sweep speed obtained by TIME/DIV,in other words, the reading is 1/10 of the sweep time indicated.



㉚ **CH1 ALT MAG** (only for model 5060)

CH1 input signal is displayed alternately by each single sweep of ×1(NORM)andx10(MAG)

set the wished portion of the waveform to the screen center for magnification.

The ×10 waveform, appears 3divisions below the ×1 waveform.

(4) Synchronization system

㉛ **SOUKCE select switch**

This switch is used to select the triggering signal source sweep.

INT The input signal applied to CH1or CH2becomes the triggering signal

LINE This setting is used when observing a signal tri-gering with power supply line frequency.

EXT External triggering signal applied to TRIG IN-PUT becomes the triggering signal.
This setting is used when triggering with a special independently of the vertical axis signal.

㉜ **INT TRIG select switch**

This switch is used to select the internal triggering signal source sweep.

CH1 The input signal applied to CH1 becomes the triggering signal,

CH2 The input signal applied to CH2 becomes the triggering signal.

VERT MODE For observing two waveforms, the sync signal changes alternately corresponding to the signals on CH1 and CH2 to trigger the signal.


㉝ **TRIG IN put connector**

Input terminal for use for external triggering signal of sweep


㉞ **TRIG LEVEL control**

This knob is used to decide at which portion of the wave-form should the sweep be started by setting trigger level. This knob is also enabled to switch SLOPE.
Depressed position(normal state)is for ⊕ SLOPE and PULL position(state in which the knob is protruding)is For ⊖ SLOPE

Explanation of synchronization polarity SLOVE



Explanation of synchronization level LEVEL



At the time of ⊕ SLOPE



At the time of ⊖ SLOPE

㉟ **TRIG MODE select switch**

AUTO The instrument is brought into automatically triggering sweep in which sweep is always conducted.
In the presence of triggered signal, normal triggered sweep is obtained and the waveform stands still. In the case of no signal or our of triggering, sweep line will appear automatically

NORM This setting is convenient in usual cases.
Triggered sweep is obtained and sweep is conducted only when triggering is effected ,No Sweep line will appear in the case of no signal or out of synchronization, Use this MODE when effecting synchronization to a very low frequency signal(25Hz or less).

TV(H) This setting is used when observing the entire vertical picture of television signal.

TV(H) This setting is used when observing the entire horizontal picture of television signal.

(NOTE) Both TV V and TV H synchronize only when the synchronizing signal is negative.

(5) Miscellaneous

㊱ **EXT BLANKING connector**

Input terminal for brightness modulation. It is of the DC coupling. The brightness is reduced with a positive signal and increases with a negative signal.

㊲ **CAL 0.5V tip**

Output terminal of calibration square wave of about 1 kHz and 0.5v .It has a tip terminal. It is used to calibrate the probe combination.

㊳ **GND terminal**

Earth terminal of the oscilloscope

5. HOW TO PRODUCE THE BRIGHT LINE

Insert the plug of the power cord on the rear panel into the power supply wall socket and set the controls as follows.

POWER	OFF
INTEN	Counterclockwise to the full
FOCUS	Midrange
AC-GND -DC	GND
▲▼ POSITION	Midrange (the knob is in the de-pressed)
V.MODE	CH1
TRIG	AUTO
TRIG SOURCE	INT
INT TRIG	CH1
TIME/DIV	0.5ms/div
◀▶ POSITION	Midrange

Set all the levers of the switches to the upper side

After ending all the setting mentioned above, turn ON the POWER and, 15 second later, rotate the INTEN knob clockwise. Then the sweep bright line will appear .If the observation is to be started immediately set the FOCUS control at a point where the bright line is sharpest

if the instrument is not used with the power supply turned on rotate the INTENSITY counterclockwise to reduce the brightness and also blur the FOCUS.

NOTE

For usual observation, leave the following non-calibrating function section set to "CAL"position.

VARIABLE	Rotate in the direction of arrow. In this case the VOLTS/DIV is calibrated to its indicating value.
SWP VAR	Leave the knob in depressed state. In this case the TIME/DIV is calibrated to its indicating value

Align the bright line with the horizontal scale line at the center of the screen by opetatingCH1 POSITION In some cases the bright line ,may be oblique to scale slightly by the effect of earth magnetism. In this case, bring the bright line until it lies on the horizontal scale line at the center of the screen by properly adjusting the semi-fixed variable resistor TRACE ROTATION on the front panel.

CENERAL MEASUREMENT

(1)In the case of observing a single waveform.

Use CH1 or CH2 when not observing the phase difference between two waveforms or when engaging in a operation other than X-Y operation. Make the following setting when using CH1

MODE Switch of Vertical defection system	CH1
MODES witch of TRIG	AUTO
TRIG SOURCE	INT
INT TRIG	CH1

Under these setting almost all the repetitive signals of about 25 Hz or more applied to CH1 can be synchronized and observed by adjusting to TRIG LEVEL, since the MODE of horizontal axis is at AUTO position, the bright line appears even when no signal is present or when input the coupling switch is at GND position, This means that the measurement of DC voltage can be measured. The following switching is needed when observing low frequency signals about 25 Hz or less.

MODE of TRIG NORM

Synchronization can be effected by operating LEVER knob under this setting.

When using only CH2, use the instrument after making the following settings.

MODE Switch of Vertical Axis	CH2
TRIG SOURCE	INT
INT TRIG	CH2

(2) When observing two waveforms

Observation of two waveforms can be made easily by setting the MODE switch of vertical axis to ALT or CHOP. When observing two waveforms of high repetition frequencies set the MODE switch to ALT and, in the case of low frequencies, set it to CHOP.

When measuring the phase difference, measure after effecting synchronization with leading phase signal.

6. METHOD FOR CONNECTING SIGNALS

The first step of measurement is introduce the signal desired to measure the oscilloscope properly. Do it with utmost care.

(1) When using a probe


Use the attached probe when measuring a high frequency wave with high accuracy.

It should be noted, however. that since the input signal is attenuated by this probe to 1/10before it is input to the oscilloscope the use of the probe is disadvantageous for low signals, and that at the same time the measuring range is extended by that amount for high signals.

<CAUTIONS>

- Do not apply a signal which exceed 400V(DC+peak AC at 1 kHz).
- Bring the grounding point of the earth lead wire of the probe close to the point to be measured when measuring a rapid rising signal or a high frequency signal. Long earth lead wire may cause waveform distortions such as ringing and overshoot.

Connection of earth lead wire



(a)A good example (b)A bad example

For better measurement it is required to use an earth attachment available at option.
Multiply the reading of VOLTS/DIV by10
For example ,if the VOLTS/DIV is 50mV/DIV,the read the wave form as

500mV/div × 10=500mV/div

To avoid measurement error, put the probe in the following correction state and check it before measurement without fail.
Connect the tip of the probe to the output terminal CAL 0.5Vof 1kHz calibration square wave voltage.
When this correction capacity value is at optimum the waveform takes the shape as shown in Fig(a)as follows.

If the waveform is as shown in Fig.(b) or Fig.(c), rotate the semifixed adjusting screw on the matching box of the probe by using a screwdriver until the optimum state is obtained.





(a)Optimum (b) Capacity too small (c) Capacity too large

(2) At time of direct connection

when connecting a signal directly to the oscilloscope not using the attached probe (10.1),pay attention to the following points in order to minimize the measurement error.

When performing observation using a bare lead wire ,no trouble occurs of the circuit to be measured is of low impedance and high level.

However, note that, in most cases, measurement error. May be caused by static stray coupling with other circuit and power line.

This measurement error cannot be ignored even in low frequency region.

In general, it is safe to avoid measuring with noon-shielded connecting wire. When using a shielding wire connect one end of the shied to the earth terminal of the oscilloscope and the other end to the grounding of the circuit to be measured. It is desirable to use a coaxial cable with BNC type connector.

The following cautions must be observed when performing a wide band measurement, It is necessary to terminate with the characteristic impedance of the cable when measuring a rapid rising waveform or a high frequency wave.

Especially when using a long cable, the absence of a terminating resistor will necessarily lead to a measurement error derived from ringing phenomenon. Some measuring circuits require a terminating resistor equal to the characteristic impedance of the cable also on the measurement terminal side.

BNC type terminating resistor(50 Ω)is conveniently used for this purpose.

In order to perform measurement with the measuring circuit put in proper operating state it is sometimes necessary to terminate the cable with an impedance which corresponds to the circuit to be measured.

The stray capacity to the shield wire must be taken into account when performing measurement with a long shied wire. Since the shield wire normally in use has

a capacity about 100 pF per meter, its effect on the circuit to be measured cannot be ignored Use a probe to minimize the effection the circuit.
When the length of the shield wire used or when the length of the non-terminated cable reaches 1/4wave length or its multiples within the band of V-5060/V-552 type(1/4 wavelength is about 1.5 meter (about 1.2 meter) when using a coaxial cable at 60MHz (50MHz),oscillation may be caused near 5mV/DIV range.
This is caused by the resonance between the externally connected high-Q inductance and the input capacity and can be avoided by reducing the Q.
Connect the cable or shield wire to the input connector by way of a serially connected 100 Ω to 1 kΩ resistor, or perform measurement at/ other VOLTS/DIV range.

(3) When observing waveform with X-Y

Set the TIME/DIV switch to X-Y. Then the instrument works as an X-Y oscilloscope.
Each input is applied to the instrument as follows:

X-axis signal (horizontal axis signal)	CH1 INPUT
Y-axis signal (vertical axis signal)	CH2 INPUT


In this case leave the horizontal axis magnification switch (PULL-MAG x 10knob) at depressed position.

7. MEASURING PROCEDURE

The first things to do are as follows:
Bring the brightness and FOCUS at optimum positions for easy read out.
Display the waveform as large as possible to minimize the read error.
Check the capacity correction when using a probe. (Refer to Paragraph (1) When using a probe of Section 6. Method for connecting signals for the method for correcting capacity.)

(1) DC voltage measurement

Set input coupling to GND and decide the zero level properly.
Set VOLTS/DIV appropriately and set AC-GND-DC to DC. Since the bright line shifts here by the amount of DC voltage, the DC voltage of the signal can be obtained by multiplying the shift width by the indicated value of VOLTS/DIV. When VOLTS/DIV is 50 mV/DIV, then 50 mV/div x 4.2 = 210 mV/ However, if the probe (10:1) is in use, the true value of the signal becomes 10 times the value, or 50 mV/div x 4.2 x 10 = 2.1V)




(2) AC voltage measurement

The same as paragraph 7 (1), "DC voltage measurement", but here those is no need of matching the zero level with the scale line. Move the zero level at will to a position easy to observe.

In the drawing is follows, VOLTS/DIV is 1V/DIV, 1V/div x 5 = 5 Vp-p (50 Vp-p at time using the probe (10:1).

When magnifying and observing a small-amplitude signal, superimposing on a high DC voltage, set input coupling of AC. The AC voltage is cut off and AC voltage can be observed by increasing sensitivity.



(3) Measurement of frequency and period

This will be explained taking the drawing at follows as an example.

One period covers the time A and time B, which are separated from each other by 2.0div on the screen.


When the sweep time is 1 ms/DIV, the period is given by

$$1 \text{ ms/div} \times 2.0 = 2.0\text{ms}$$
$$= 2.0 \times 10^{-3} \text{ s}$$

Accordingly.the frequency is


$$1/(2.0 \times 10^{-3}) = 500\text{Hz}$$

(However, when the knob MAG×10 is at pulled out position, TIME/DIV must be converted to1/10 since the sweep is magnified.)



(4) Measurement of time difference

Triggering signal source"SOURCE"is selected as offering reference signal when measuring the time difference between two signals. Assume that pulse trains as shown in(a). Then(b)shows the case when CH1 is take as the triggering signal source and (c)the case where CH2 is taken.




This means that Ch1 is used as the triggering signal when investigating the length of time by which the signal of CH2 is delayed from the signal of CH1.CH2is used in the reversed case. In other words ,the signal leading in phase is selected as the triggering signal source.


If this process is reversed ,the portion to be measured may sometimes not appear on the screen .Thereafter. equalize the amplitudes of the two signals appearing on the screen or superimpose one on another.

Read the time difference by the interval between 50% amplitude points of the two signals.

Sometimes the superimposing method is more convenient from the point of view of procedure.



(a) Equal amplitude measuring method



(b) Superposition measuring method

<CAUTIONS>

Since the pulsed wave contains many high-frequency wave components (higher harmonics)dependingon its width width or

Period, pay the same attention as give to high frequency signals when handling it. Accordingly ,use a probe or coaxial cable and shorten the earth lead wire as possible.

(5)Measuremnt of rise (fell)time

To measure the rise time pay attemtion not only to the abovementioned items but also to measurement error. The following relationship exists between the rise time Trx of the waveform to be measured, the rise time Trs of oscilloscope, and the rise time Tro displayed on the screen.

$$Tro=\sqrt{Trx^2 + Trs^2}$$

When the rise of the pulse going to be measured is sufficiently. longer than the rise time of the oscilloscope (7ns in our case).the effect of the rise time of the oscilloscope on the measurement can be neglected ,However ,if both are close to each other, measurement error may be caused.

The true rise time is given by

$$Tro=\sqrt{Tro^2 + Trs^2}$$

Moreover, in general, in a circuit free from waveform distortion such as overshoot and sag, the following relationship is established between frequency band and rise time.


$$f_c \times t_r = 0.35$$

Where, f_c: Frequency band(Hz)
tr :Rise time (S)

The rise time and fall time are determined by the time elapsed between the 10% to90% values of pulse width This oscilloscope is provided with graduations for 0%, 10%, 90%, and 100% on the screen, which facilitate measurement.

(6) Synchronization of complexed waveform

In the case shown in the fig. (a) below where waveforms greatly different in amplitude alternate. the waveform is doubled if the trigger level is not set properly .In the case where the trigger level is selected as Y line two waveforms, one starting with A and advaneing to B, C, D, E, F, . . . , And the other starting with E and advaneing to F, G, H, I . . . , Will appear alternately, on the screen. They will be doubled as shown in Fig(b), for which no synchronization can be taken.
In such a case, rotate LEVEL clockwise until the trigger level comes to Y line ,Then the waveform on the screen becomes the one is shown in Fig. (c)above which start with Band advances to C, E, F, . . . and which allows synchronization



Synchronization of complexed waveform

(7)synchronization of observing two waveforms

① When two signals of CH1 and CH2 have same frequencies or the frequencies of an integral number or the frequencies in a relation of a specific time difference, the INT TRIG switch selects either CH1or CH2 as a reference signal. CH1 position selects CH1 signal as a reference, and CH2 position selects CH2 signal.

② For an observation of signals fo different frequencies set the INT TRIG switch to the VERT MODE The sync signal switches at each alternation of channels ,and the waveform of the each channel is stably triggered.

SELECTION OF A TRIGGER SOURCE ON THE VERT MODE

A. Trigger signal is obtained the following steps

1)set the SOURCE switch (31) to INT.

2)set the INT TRIG switch (32) to VERT MODE

3)select the MODE switch

Table 1.

Relation of trigger signal sources and switches

SOURCE		INT			LINE	EXT
INTTRIG		CH1	CH2	VERT MODE	Line	External
VMODE	CH1	CH1	CH2	CH1		
	CH2	CH1	CH2	CH2		
	ALT	CH1	CH2	CH1 CH 2(ALT)		
	CHOP	CH1	CH2	ADD		
	ADD	CH1	CH2	ADD		

When the SOURCE switch to INT, TRIG switch to VERT MODE, and MODE switch to ALT, the input signals applied to CH1 and CH2 become trigger source alternatively at each sweep. Consequently, even for an observation of two waveforms of different frequencies, the waveform of the each channel is stably triggered.

In this case ,the signal should be applied to both CH1 and CH2 and the two signals have the same level portion in excess of the rated amplitude each other. There should be a common portion of levels available that is above the rated amplitude of CH1 and CH2.

when a sinewave is applied toCH1, and a square wave is applied to CH2, "A"s in Fig. 1 are the levels possible for synchronization.




Fig.1

In order to expand the synchronization range , AC coupling is applied to the Ch2side.

When either the Ch1or CH2 input signal is smaller as shwonin Fig.2, adjust the VOLTS/DIV switches (13) and (14) to obtain sufficient amplitudes.

The VERT MODE triggering requires 1.5 div mire than the amplitude required for an observation of ch1or CH2.




Fig.2

The CERT MODE triggering is not possible when the signal is applied to only one channel as illustrated on Fig.3.





Fig3

Caution: Do not use the INT TRIG to the VERT MODE when PULL×5 GAIN ⑮ and/or ⑯ are in the pulled out position(×5 GAIN mode)

ALTERNATE TRIGGER


Jittering wave as shown below may appear on the screen when a gently-sloping signal is displayed by approximate 10 cycles or less .with the VERT MODE for the INT TRIG switch and the ALR position for the MODE select witch. For detailed and clear observation of each signal, set the MODE select switch Ch1or Ch2





(8)How to use TV exclusive synchronization

① On the image waveform of TV

In the work concerned with TV, complexed signals ontaining video signal, blanking pedestal sitgnal, and synchronizing signal are often measured, However, since the waveform is complexed, a special circuit is needed to effect a stable synchronization with vertical waveform.




②Difference in the circuits

	Exclusive circuit for conventional oscillograph		Exclusive circuit for this instrumet (principle drawing)
	General circuit	simple synchronizing circuit	TV exclusive synchronizing separator circuit
Circuits	Video signal To triggetr circuit 	To trigger circuit 	Vc  To trigger circuit
	Hard to synchronize, because video signal applied directly as trigger signal.	Synchronization is more easily effected than in the circuit shown at left ,because the signal is integrated to remove high frequency components.	Stable synchronization is on-tained since SYNC pulse is picked up, amplified, and then integrated to remove high fre-quency components.


③Operation

To observe vertical signal



TRIG MODE: TV-V


To observe horizontal signal



TRIG MODE: TV-H

(NOTE) This oscilloscope synchronizes with only (-)synchronizing signal.

(REFERENCE)



(a)Example of(-) synchronizing signal

(b)Example of(+) synchronizing signal

(9) DC OFFSET

The oscilloscope provides the Dc offset voltage display or ±1 to ±100V accordingto the range
An output terminal for voltage reading
(Except : ×5GAIN, out of CAL)




Fig.7-1 Diagram of DC OFFSET

see Fig ,7-1for the function of the DD OFFSET
There is following relation among the differential amplifier output voltage e_o, the input voltages ei and E_2

$e_o=A(e_i-E2)$

$e_i=E_{DC}+e_{AC}$

(where E_{DC} is a DC input and e_{AC} is an AC input)

set the controls so that $E_2=E_{DC}$ is attained. Then $e_o=Axe_{AC}$ is obtained.

The DC component can be removed for an observation

OFFSET voltage ranges

VOLTS/DIV	DC OFFSET voltage
5mV/div-50mV/div	More than±1V(x1)
0.1V/div-0.5V/div	more than±10V(x10)
1V/div-5V/div	more than±100V(x100)

When measured with a DMM connection, multiply the DMM reading by the multiplier written above in indentation

Read next paragraph for the detailed explanation of measurement with a DMM connection.

(10)Measurement by the DC OFFSET function

In order to readout the voltage level digitally, connect a DMM (digital multimeter) to the DC OFFSET output terminal and set the oscillosecope to the DC OFFSET mode. see Fig.7-2

1)Measuting DC component

Ailgn level ⓑ with the center graticule and read the DMM digital value (+2V should be displayed.)




Fig.7-2 Transistor amplifier

2)Measuring AC component

When the above is performed with an input sensitivity of 50mV then 4 div of amplitude must be present on the oscilloscope screen.

The AC component is observable on the oscilloscope screen where as the DC component us measurable on a DMM, and there is no need for any complicated switching as such needed for conventional oscilloscopes.

More over the DC OFFSET function provides readings of peak to peak value(p-p)of e_o on DMM.

Align level ⓐ with the center graticule line on the screen ,read the DMM value and name it V(a), then ailgn level Ⓒ with the graticule line and readout V (c).Thep-p

the difference between V(a) and V(c), and the value can be read digitally on the DMM.
As explained above, the DC OFFSET function with DC OFFSET output terminal offers an improved operational convenience and a highly accurate measurement in detailed waveform portions.

8. ADJUSTMENTS

The ATT balance of the vertical axis can be made easily.

① Set the input coupling switches of CH1 and CH2 to GND and set the TRIG MODE to AUTO. Then position the bright line to the center.
② Turn the VOLTS/DIV switch to 5mV-10mV and adjust so that the bright line does not move.
(DC BAL ㉔ ㉕)

9. MAINTENANCE

1) Since semiconductors ,precision components, etc.are employed in this oscilloscope, use at most care for oparation and storage.
2) Clean the scale with soft tissue periodically
3) Side panel can be removed with screws
4) Store this oscilloscope in the ambient temperature from -10℃ to +50℃.

10. SPECIFICATIONS

CRT

Type

Large6"screen with internal graticule.
Approximate 12kv acceleration potential.

Phosphor

P31 standard

Graticule

8×10div(div=10mm)
Internal graticule

Focussing
possible (with automatic focus correction circuit)

Trace rotation	present
Brightness adjustment	possible
Scale illumination	variable

Z-AXIS INPUT(INTESITY MODULATION)

DC-coupled ,positive-going signal decreases intensity: 5Vp-p signal causes noticeable modulation at normal

Intensity: DC to 2MHz
Input impedance 33kohm(typ.)
Maximum input voltage 30v(DC+peak AC)

VERTCAL DEFLECTION SYSTEM(2identical channels)

Bandwidth and rise time

DC to at least 60MHz (50MHz) and rice time 7ns [17.5 ns] or less. DC to at least 7 MHz and rise time 50ns or less .at magnifier extends, The AC coupled lower -3dB point is 10Hz or less.

Deflection factor

5mV/div to 5V /div in10 calibrated steps in a 1-2-5 sequence. Uncailbrarted continuous control extends deflection factor to at least 12.5 Volts per division in the 5 Volts/div position x5 magnifier increases sensitivity of each deflection factor setting to 1mv/div

Accuracy

±3%
Additional error for magnifier±2%

Display modes

CH1,CH2(normal or invert), Alternate, chopped (approximate 250kHz), Added

Input Impeuance

Approximately 1 MΩ in parallel with 25 pF

Maximum input voltage

300V(DC+pake AC)or 500Vp-pACat 1 kHz or less

Input coupling

AC, GND, DC
DC OFFSET
DC OFFSET VOLT OUT

HORIZONTAL DEFLECTION SUSTEM

Time base
0.2 μ s/div to 0.2s/div in 19 calibrated steps in a-1-2-5 sequence, Uncalibrated continuous control extends deflection factor to at lease 0.5 seconds per division in the 0.2sec/div position x10 mag extends maximum sweep rate to 20 ns/div [100ns/div] .

Accutacy

±3%
Additional error for magnifier ±2%

TRIGGERING SYSTEM

Trigger modes

Automatic Normal, TV(TV-H or TV-V)

Trigger source

Internal (CH1,CH2 v-MODE),Line,External

Trigger slope

+,-

TV sync polarity

TV(-)

TV-Vsensitivity :SYNC section less 1 div ir1V

AUTO low band: Approximately 25Hz

Trigger coupling

AC : 20Hz to full band width

External trigger input impedance

Approximately 1MΩ In parallel with 30pF

Maximum input voltage

300V(DC+AC peak)

ALT MAG present

X-Y OPERATION (CH1; Horiz,CH2; vert,

Deflection factor

same as Vertical deflection

X-bandwidth

DC to at least 500kHz

Phase error

3℃ or less from DC to 50kHz

CALIBRATOR

An approximate 1kHz frequency 0.5V（±3%) square wave.

SIGNAL OUTPUT

CH1 VERT SIGNAL OUTPUT

Output voltage is at least 20mV/div into a 50 ohm load.
Bandwidth is 50Hz to at least 5MHz

POWER SUPPLY

VOLTAGE(50Hz)	FUSE
220V(198-242V)	0.5A

Power supply frequency :50Hz
Power consumption: Approx. 35W [30W]
Max.40W at 120V 60Hz


ENVIRONMENT


Limit of operation temperature	0-+40℃
Limit of operation humidity	35-85%
Rated range of use temperature	+10-+35o℃
Rated range of use humidity	45-85%
storage and transport temperature	-10℃-+50℃


DIMENSION AND WEIGHT

Approx.310(w)×130(H)×370(D)mm
(12.4(W)×5.2(H)×14.8(D)inch)
Approx.6.5kg (14.61bs) [6kg(13.51bs)]

11. EXTERNAL VIEW







V-552 V-5060 Unit:mm

MUMBAI

TEST CERTIFICATE

DIGITAL ANALOG STORAGE OSCILLOSCOPE

This Test Certificate warrantees that the product has been inspected and tested in accordance with the published specifications.

The instrument has been calibrated by using equipment which has already been calibrated to standards traceable to national standards.

MODEL NO. _____

SERIAL NO. _____

DATE: _____

ISO 9001REGISTERED

QC
KUSAM-MECO
PASS

WARRANTY

Each “KUSAM-MECO” product is warranted to be free from defects in material and workmanship under normal use & service. The warranty period is one year (12 months) and begins from the date of despatch of goods. In case any defect occurs in functioning of the instrument, under proper use, within the guarantee period, the same will be rectified by us free of charges, provided the to and fro freight charges are borne by you.

This warranty extends only to the original buyer or end-user customer of a “KUSAM-MECO” authorized dealer.

This warranty does not apply for damaged Ic’s, burnt PCB’s, fuses, disposable batteries, carrying case, test leads, or to any product which in “KUSAM-MECO’s” opinion, has been misused, altered, neglected, contaminated or damaged by accident or abnormal conditions of operation or handling.

“KUSAM-MECO” authorized dealer shall extend this warranty on new and unused products to end-user customers only but have no authority to extend a greater or different warranty on behalf of “KUSAM-MECO”.

“KUSAM-MECO’s” warranty obligation is limited, at option, free of charge repair, or replacement of a defective product which is returned to a “KUSAM-MECO” authorized service center within the warranty period.

THIS WARRANTY IS BUYER’S SOLE AND EXCLUSIVE REMEDY AND IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESS OR IMPLIED, INCLUDING BUT NOT LIMITED TO ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. “KUSAM-MECO” SHALL NOT BE LIABLE FOR ANY SPECIAL, INDIRECT, INCIDENTAL OR CONSEQUENTIAL DAMAGES OR LOSSES, INCLUDING LOSS OF DATA, ARISING FROM ANY CAUSE WHATSOEVER.

All transaction are subject to Mumbai Jurisdiction.

KUSAM-MECO

17, Bharat Industrial Estate, T. J. Road, Sewree (W),
Mumbai-400015. INDIA
Sales Direct: 24156638 **Tel.:**(022)2412 4540, 2418 1649 **Fax:**2414 9659
E-mail : kusam_meco@vsnl.net,
Website : www.kusamelectrical.com, www.kusam-meco.co.in